

Method Detection/Reporting Limits Method TO-9A (Dioxin/Furan) Air

TO9_Air_03_03.xls



Analyte	Method Detection Limit (ng/S)	Method Reporting Limit (ng/S)	LCS Criteria Lower Control Limit (%)	LCS Criteria Upper Control Limits (%)	% RPD Limit
2,3,7,8-TCDF	0.0017	0.010	70	130	30
2,3,7,8-TCDD	0.0015	0.010	70	130	30
1,2,3,7,8-PeCDF	0.0023	0.050	70	130	30
2,3,4,7,8-PeCDF	0.0016	0.050	70	130	30
1,2,3,7,8-PeCDD	0.0016	0.050	70	130	30
1,2,3,4,7,8-HxCDF	0.0014	0.050	70	130	30
1,2,3,6,7,8-HxCDF	0.0016	0.050	70	130	30
2,3,4,6,7,8-HxCDF	0.0015	0.050	70	130	30
1,2,3,7,8,9-HxCDF	0.0012	0.050	70	130	30
1,2,3,4,7,8-HxCDD	0.0038	0.050	70	130	30
1,2,3,6,7,8-HxCDD	0.0024	0.050	70	130	30
1,2,3,7,8,9-HxCDD	0.0036	0.050	70	130	30
1,2,3,4,6,7,8-HpCDF	0.0016	0.050	70	130	30
1,2,3,4,7,8,9-HpCDF	0.0033	0.050	70	130	30
1,2,3,4,6,7,8-HpCDD	0.0010	0.050	70	130	30
OCDF	0.0057	0.10	70	130	30
OCDD	0.026	0.10	70	130	30

Pace Analytical Services, Inc. is a NELAC accredited laboratory organization and meets NELAC testing standards. Use of the NELAC logo however does not insure that Pace is currently accredited for the specific method indicated. For current Pace accreditation information consult your Pace Project Manager.